



Stack Emissions Proficiency Testing Scheme (SEPTS)

Presentation of Results

Round 2025

EffecTech is accredited by the United Kingdom Accreditation Service (UKAS) to provide this Proficiency Testing Scheme in accordance with the requirements of ISO/IEC 17043: 2010

Dove House Dove Fields Uttoxeter Staffordshire ST14 8HU United Kingdom

T +44 (0)1889 569229 F +44 (0)1889 569220 www.effectech.co.uk

Document Ref.: 25/0642-0648/R Revision: 1

Document Date: 29 July 2025 Author(s): Joey Walker

Table of Contents

Tab	le of Contents	2
Rev	risions History	3
Stat	tement of Confidentiality	4
1.	Introduction	5
2.	Mixture preparation and reference value assignment	6
2.1	Procedure	6
	Preparation of mixture batches	6
	Mixture calibration	6
	Reference mixture traceability	6
	Homogeneity assessment	6
	Reference value assignment	7
	Stability statement	7
2.2	Assigned reference values	8
3.	Results	9
3.1	Reported results	9
3.2	Measures of performance	11
	z-score	11
	E _n number	12
3.3	Evaluation of results	13
Ann	nex A - Detailed results by measurand	15
Anr	nex B - Converter efficiency	32

Revisions History

Issue	Date	Author(s)	Comments
1	29 July 2025	Joey Walker	Draft report (for comment)

25/0642-0648/R Page **3** of **33**

Statement of Confidentiality

EffecTech keeps all data regarding the performance of individual participants strictly confidential. Results and performance data are protected, stored and backed up on storage network disks and folders to which access is restricted to the scheme coordinator and the technical authority only.

The relationship between results and the laboratories that submitted them will never be disclosed. The laboratory alone is granted access to its performance through the assigned participant code and through issue of a confidential Certificate of Participation.

Checked by

Approved by

Steve Price

Scheme Coordinator

S. Price

Joey Walker

Technical Authority

25/0642-0648/R Page **4** of **33**

1. Introduction

EffecTech provides and organises the Stack Emissions Proficiency Testing Scheme (SEPTS). This report presents data on the reference mixtures in cylinders and the results of the participants for Round 2025 (June - July 2025).

The SEPTS scheme provides an objective way of assessing the performance of each participant by a series of annual inter-laboratory comparisons. The scheme is aimed at laboratories/testing organisations working in the field of continuous emissions monitoring (CEM) of stationary sources often in waste incineration or large combustion plant processes.

In this round participants were given the opportunity of analysing up to eight (8) different measurands in seven (7) gas mixtures. The composition range of each measurand in each mixture is shown in the tables below.

Table 1: Composition range by gas mixture type

measurand	range
sulphur dioxide (SO ₂)	10 to 200 μmol/mol
in nitrogen	10 το 200 μποι/ποι
propane (C₃H8)	5 to 50 μmol/mol
in 10% oxygen / nitrogen	3 το 30 μποι/ ποι
nitric oxide (NO)	10 to 500 μmol/mol
in nitrogen	10 to 300 μmol/mol
carbon monoxide (CO)	20 to 500 μmol/mol
in nitrogen	
oxygen (O ₂)	2 to 14 %mol/mol
in nitrogen	, .
carbon dioxide (CO ₂)	1 to 10 %mol/mol
in nitrogen	,
nitric oxide (NO) and	20 to 400 μmol/mol
nitrogen oxides (NO _x)	25 to 500 μmol/mol
in nitrogen	

Note: all units used in this report are in the SI unit of amount of substance fraction (mol/mol) or in metric prefixes thereof. $500 \,\mu$ mol/mol is equivalent to $500 \,x \, 10^{-6} \,m$ ol/mol

10 %mol/mol is equivalent to 10 dmol/mol is equivalent to 10×10^{-2} mol/mol

Gas mixture preparation, reference value assignment and the assessment of participants' results are all carried out by designated operators and approved signatories within EffecTech and in accordance with our ISO/IEC 17043 accredited processes.

In addition, all logistics management and preparation of shipping documentation is also carried out by designated approved personnel within EffecTech. All shipping, freight forwarding and proficiency testing item distribution is supplied by specialist transport providers.

A total of twenty-seven (27) laboratories signed up to participate in this round. Twenty-seven (27) laboratories to whom items were distributed, submitted results for one or more of the measurands assessed in the scheme.

25/0642-0648/R Page **5** of **33**

2. Mixture preparation and reference value assignment

2.1 Procedure

Preparation of mixture batches

For each mixture type, a single large volume parent mixture was prepared by a gravimetric method in accordance with ISO 6142-1. A batch of mixtures of each type was then produced by decanting the parent mixture into a batch of lower volume pre-prepared and evacuated daughter cylinders. The parent mixture and daughter mixtures were then calibrated.

Mixture calibration

All parent mixtures were calibrated using a two-point calibration design with bracketing (TPC), with the exception of the carbon dioxide measurand which was calibrated using a single-point through origin calibration (SPO). Both of these calibration methods are in accordance with ISO 12963 for which EffecTech is accredited to ISO/IEC 17025 and ISO 17034 by UKAS.

Every single decant mixture was calibrated by a single point exact matching technique (SPEM) also in accordance with ISO 12963 by the comparison of the decant mixture with its nominally identical parent mixture. A selective batch calibration technique was not used. All mixtures despatched to participants were calibrated individually to provide ultimate assurance in the assigned reference values.

The uncertainty on amount fraction of each measurand in the mixtures resulting from this calibration is termed the characterisation uncertainty, u_{char} (ISO 33405).

All calibrations are performed in accordance with EffecTech's ISO/IEC 17025 accredited calibration methods (in-house methods TM014, TM025 or TM026). These can be found in our scope of accreditation published on the United Kingdom Accreditation Service (UKAS) website (www.ukas.com).

Reference mixture traceability

An analytical comparison method is used for the calibration of all mixtures in this scheme. In-house primary reference gas mixtures (PRGM) are used for calibration which are traceable by verification to the National Physical Laboratory (NPL, UK) or to the Van Swinden Laboratorium (VSL, NL). Parent mixtures are calibrated either by direct comparison with PRGMs (SPO) or, where diluted, with reference gas mixtures generated dynamically in accordance with ISO 6145-7 (TPC). This process ensured that the values assigned to the mixtures in this scheme are metrologically traceable to international standards, through an unbroken chain of comparisons, and ultimately to the amount of substance (mole) defined in the SI (International System of Units).

Homogeneity assessment

Statistical analysis of the spread of reference values obtained for each batch of mixtures (derived through calibration above) is used to assess the homogeneity between the set of decant mixtures to be distributed to each participant. The dispersion of the amount fraction of each component due to batch inhomogeneity is known as the between-bottle standard deviation (s_{bb}). The uncertainty arising from this is the between-bottle uncertainty (u_{bb}). The statistical procedure used for the determination of u_{bb} = s_{bb} can be found in ISO 33405.

25/0642-0648/R Page **6** of **33**

The uncertainty associated with within-bottle heterogeneity u_{wb} is assumed to be zero; EffecTech has conducted numerous measurements and intercomparisons that demonstrate that for well mixed gas mixtures, samples used for analysis are homogenous with the gas mixture within the cylinder. Hence the standard uncertainty associated with heterogeneity $u_{hom} = u_{bb}$. This uncertainty should be less than or equal to the characterisation uncertainty, u_{char} , in order to accept the batch. This condition was met for all components in all mixtures produced for all participants in this round.

Reference value assignment

For all measurands, each component was assigned a reference value, x_{ref} , calculated from the average (simple arithmetic mean) of those determined in the calibration stage (see section above).

The initial combined uncertainty determined for each reference value was calculated from the equation below (ISO 33405 – section 10.2).

$$u_c^2 = u_{char}^2 + u_{hom}^2$$

This combined uncertainty, u_c, is dominated in all cases by the calibration uncertainty, u_{char}

Following this calculation, the expanded uncertainty, $k \cdot u_c$, (k=2), was compared to the Calibration and Measurement Capability (CMC) for which EffecTech is accredited to ISO/IEC 17025. If U_{CMC} (k=2) was greater than $k \cdot u_c$ (k=2) then the uncertainty on the reference value was assigned to that stated in our published CMC in accordance with accepted practice such that

$$U_{ref} = max (U_{CMC}, 2u_c)$$

The use of a coverage factor of k=2 in the assignment of U_{ref} provides a level of confidence of approximately 95%.

The individual calibration data for each suite of decant mixtures is not shown in this report. However, this data is available to all participants on request from EffecTech.

Stability statement

Over several years EffecTech has built up a history of intercomparisons of mixture types featured in this PT scheme. Data from these intercomparisons show clearly that all mixtures remain stable within their stated uncertainty for a minimum of 12 months.

Hence, the stability of each mixture is guaranteed for a period of 12 months. Within this time period there is no additional uncertainty ascribed to the reference values resulting from the long or short term stability of the mixtures. This is valid providing the mixtures are not used beyond this stability period.

The majority of these mixtures will be stable (within their stated uncertainty) for considerably longer but this period has not been determined.

25/0642-0648/R Page **7** of **33**

2.2 Assigned reference values

The table below show the reference values assigned to the measurands in the mixtures in cylinders distributed to participants of this scheme.

Table 2: Reference values assigned following batch homogeneity assessment

measurand	units	Xref	$U(x_{ref})$	u_c / %	u _{char} / %	u _{bb} / %
sulphur dioxide	μmol/mol	115.6	1.2	0.51	0.50	0.10
propane	μmol/mol	27.14	0.26	0.35	0.34	0.082
nitric oxide	μmol/mol	82.45	0.46	0.23	0.22	0.049
carbon monoxide	μmol/mol	235.8	1.9	0.35	0.35	0.035
oxygen	%mol/mol	6.827	0.037	0.056	0.056	0.0030
carbon dioxide	%mol/mol	6.164	0.025	0.21	0.20	0.048
nitric oxide (NO/NO2 mix)	μmol/mol	86.57	0.48	0.23	0.22	0.053
nitrogen oxides (NO/NO2 mix)	μmol/mol	97.77	0.69	0.36	0.34	0.11

25/0642-0648/R Page **8** of **33**

3. Results

3.1 Reported results

There were twenty-seven (27) laboratories/organisations signed up for participation in this round of the scheme. Consignments containing up to seven (7) different mixture types were shipped to those participating.

The tables below show participation and whether results were submitted for the mixtures shipped.

Table 3: Participant laboratories and reported results

Participant id	sulphur dioxide		propane		nitric o	xide	carbon monoxide	
·	participation	results	participation	results	participation	results	participation	results
P01	✓	✓	✓	✓			✓	✓
P02								
P03							✓	✓
P04								
P05					✓	✓		
P06	✓	✓	✓	\checkmark			✓	✓
P07	✓	✓	✓	✓	✓	✓	✓	✓
P08	✓	✓	✓	✓	✓	✓	✓	✓
P09	✓	✓	✓	✓	✓	✓	✓	✓
P10							✓	×
P11	✓	✓	✓	\checkmark			✓	✓
P12	✓	✓	✓	✓	✓	✓	✓	✓
P13			✓	\checkmark	✓	\checkmark	✓	✓
P14			✓	✓	✓	✓		
P15	✓	✓	✓	✓			✓	✓
P16			✓	✓				
P17	✓	✓	✓	✓	✓	✓	✓	✓
P18	✓	✓	✓	✓	✓	✓	✓	✓
P19								
P20								
P21	✓	✓	✓	\checkmark	✓	✓	✓	✓
P22	✓	✓	✓	✓	✓	✓	✓	✓
P23	✓	✓	✓	✓	✓	✓	✓	✓
P24	✓	✓	✓	✓	✓	✓	✓	✓
P25	✓	✓			✓	✓	✓	✓
P26	✓	✓	✓	✓	✓	✓	✓	✓
P27	✓	✓	✓	✓			✓	✓

25/0642-0648/R Page **9** of **33**

Participant id	oxygen		carbon dioxide		nitric oxide (NO/NO2 mix)		nitrogen oxides (NO/NO2 mix)	
	participation	results	participation	results	participation	results	participation	results
P01	✓	✓	✓	✓	✓	✓	✓	✓
P02					✓	✓	✓	✓
P03	✓	✓	✓	✓	✓	✓	✓	✓
P04	✓	✓	✓	✓				
P05	✓	✓	✓	✓				
P06	✓	\checkmark	✓	✓	✓	✓	✓	✓
P07	✓	\checkmark	✓	\checkmark	✓	\checkmark	✓	✓
P08	✓	\checkmark	✓	\checkmark	✓	\checkmark	✓	\checkmark
P09	✓	\checkmark	✓	\checkmark	✓	✓	✓	\checkmark
P10	✓	\checkmark	✓	\checkmark	✓	×	✓	×
P11	✓	\checkmark			✓	\checkmark	✓	✓
P12	✓	\checkmark	✓	\checkmark	✓	\checkmark	✓	\checkmark
P13	✓	\checkmark	✓	\checkmark	✓	\checkmark	✓	✓
P14	✓	\checkmark						
P15	✓	\checkmark	✓	\checkmark	✓	✓	✓	\checkmark
P16								
P17	✓	\checkmark	✓	\checkmark	✓	✓	✓	\checkmark
P18	✓	\checkmark						
P19	✓	\checkmark	✓	\checkmark				
P20	✓	✓	✓	×				
P21	✓	\checkmark	✓	\checkmark	✓	✓	✓	✓
P22	✓	✓	✓	✓				
P23	✓	×	✓	✓	✓	✓	✓	✓
P24	✓	\checkmark	✓	✓	✓	✓	✓	✓
P25	✓	\checkmark	✓	\checkmark	✓	✓	✓	✓
P26	\checkmark	\checkmark	✓	\checkmark	✓	✓	✓	\checkmark
P27	✓	✓	✓	✓	✓	✓	✓	✓

To enable the calculation of E_n numbers, the laboratory is required to submit an estimate of the uncertainty placed on their measured amount fractions. The majority of participants submitted estimates of measurement uncertainty on the measurands for which they reported a value.

25/0642-0648/R Page **10** of **33**

3.2 Measures of performance

z-score

The evaluation of performance was carried out by means of a **z**-score, which gives the relative deviation of the participant's result from the reference value.

The z-score is calculated with the following general formula

$$z = \frac{x_{meas} - x_{ref}}{\sigma} \tag{1}$$

where x_{meas} is the measured result reported by the laboratory

 x_{ref} is the assigned reference value and

 σ is the absolute standard deviation used for performance assessment (SDPA) which can be calculated from the contributions $S_{PT,rel}$ and $S_{PT,abs}$ by

$$\sigma = \frac{S_{PT,rel}}{100} \cdot x_{ref} + S_{PT,abs} \tag{2}$$

If there is concern that the estimation of the **z**-score may be biased due to the magnitude of the uncertainty of the assigned reference value in the case when $u_{ref} > 0.3\sigma$ then the use of a modified **z**'-score shall be used to evaluation performance for that component failing this condition.

The z'-score is calculated with the following general formula.

$$\mathbf{z}' = \frac{x_{meas} - x_{ref}}{\sqrt{\sigma^2 + u_{ref}^2}} \tag{3}$$

The standard deviation for performance assessment used for calculating the **z**-scores has been fixed for all components by EffecTech and based upon a reasonable expectation of the performance capabilities that should be demonstrated by each laboratory for the direct measurement of a gas mixture in a cylinder.

These are given in the tables below.

Table 4: Standard deviation for performance assessment

measurand	S _{PT,rel}	S _{PT,abs}
sulphur dioxide	2.5 % relative	
propane	5.0 % relative	
nitric oxide	2.5 % relative	
carbon monoxide	1.5 % relative	
oxygen	1.0 % relative	0.01 %mol/mol
carbon dioxide	1.0 % relative	0.01 %mol/mol
nitric oxide (NO/NO2 mix)	2.5 % relative	
nitrogen oxides (NO/NO2 mix)	2.5 % relative	

25/0642-0648/R Page **11** of **33**

The qualification of the z-scores is given in table 5 below

Table 5: Relationship between z-score and quality of result

z -score	quality of result
z ≤ 2	satisfactory result
2 < z < 3	questionable result
$ z \ge 3$	unsatisfactory result

E_n number

In addition, an E_n number is calculated which assesses the difference in the reference and measured (reported) values relative to their respective uncertainties. The calculation of E_n numbers is dependent upon the laboratory estimates of uncertainties associated with their measurement results.

The E_n number is calculated with the following general formula

$$\boldsymbol{E_n} = \frac{x_{meas} - x_{ref}}{\sqrt{U_{meas}^2 + U_{ref}^2}} \tag{4}$$

where x_{meas} is the measured result reported by the laboratory x_{ref} the assigned reference value and U_{meas} and U_{ref} their respective uncertainties (using a coverage factor k=2)

The qualification of the E_n number is given in table 6 below

Table 6: Relationship between E_n-number and quality of result

E _n number	quality of result
$ E_n \leq 1$	satisfactory result
$ E_n > 1$	unsatisfactory result

Evaluation of the performance of a laboratory based on E_n numbers requires a reported estimate of their measurement uncertainty, U_{meas} . In addition, it is important that the reported uncertainties are in the same order of magnitude as the uncertainties on the reference values. Due to the nature of the formula used to calculate the E_n number, high reported uncertainties are much more likely to result in very low E_n numbers.

25/0642-0648/R Page **12** of **33**

3.3 Evaluation of results

The results of the evaluation of **z**-scores based upon the expectation SDPA are shown in the table below.

Table 7 - Summary of z-scores

z-scores

participant id	sulphur dioxide	propane	nitric oxide	carbon monoxide	oxygen	carbon dioxide	nitric oxide (NO/NO2 mix)	nitrogen oxides (NO/NO2 mix)
P01	0.26	-0.03		0.85	0.29	-0.38	0.42	-0.74
P02							-4.49	-8.37
P03				0.19	0.11	1.11	0.05	-0.05
P04					-0.15	1.41		
P05			0.54		-0.06	-1.49		
P06	1.13	0.14		1.25	0.71	-0.04	-0.18	-1.17
P07	0.51	-0.15	0.06	0.77	0.06	0.05	-0.01	-0.64
P08	0.27	-0.54	-0.74	0.10	-0.13	-0.31	-1.18	-1.91
P09	-0.55	-0.52	-0.64	0.49	-0.23	0.15	-0.77	-1.55
P10					-0.09	1.06		
P11	-0.31	-1.13		-2.25	-7.28		0.67	0.46
P12	0.17	-0.92	0.16	2.30	0.83	-2.30	-0.52	-1.72
P13		-2.45	0.37	0.03	0.74	-0.32	-1.80	1.49
P14		0.24	0.42		-0.73			
P15	-1.89	-1.74		0.13	2.85	-0.05	-1.89	-2.97
P16		0.10						
P17	-0.01	-0.23	0.32	-0.45	0.93	4.69	0.85	1.04
P18	-0.22	-0.40	-0.25	0.17	-0.11			
P19					0.70	0.90		
P20					-0.15			
P21	-0.41	0.15	1.09	-0.79	0.98	-2.32	0.98	0.21
P22	0.72	-0.48	0.53	-0.17	0.11	0.29		
P23	-0.12	-0.46	0.87	-0.14		-1.09	0.90	0.68
P24	0.47	-1.28	-0.58	-0.05	0.31	1.61	-2.57	-2.39
P25	2.21		-6.04	0.93	0.42	-2.01	-6.09	-5.80
P26	-0.50	0.00	1.14	-0.94	0.65	-2.21	1.04	0.42
P27	-0.28	0.28		1.33	-0.01	0.10	-1.28	-1.78

These results show a generally good performance from the pool of participants.

However, participants **P02**, **P24** and **P25** reported anomalously low measurements for the nitric oxide/nitrogen oxides PT item. Participant **P11** reported low on oxygen, while **P17** reported high for carbon dioxide.

Seventeen (17) laboratories (P01, P03, P04, P05, P06, P07, P08, P09, P10, P14, P16, P18, P19, P20, P22, P23 and P27) achieved satisfactory results for all measurands for which they reported a result.

25/0642-0648/R Page **13** of **33**

Performance based upon the E_n -numbers are given in the table below.

Table 8 - Summary of E_n -numbers

 E_n -numbers

participant id	sulphur dioxide	propane	nitric oxide	carbon monoxide	oxygen	carbon dioxide	nitric oxide (NO/NO2 mix)	nitrogen oxides (NO/NO2 mix)
P01	0.47	-0.05		1.12	0.15	-0.10	1.39	-2.17
P02							-6.03	-12.07
P03				0.03	0.01	0.13	0.01	-0.01
P04					-0.09	0.37		
P05			0.41		-0.12	-3.81		
P06	2.21	0.21		1.29	1.38	-0.11	-0.24	-1.54
P07	0.20	-0.09	0.02	0.25	0.02	0.01	0.00	-0.20
P08	0.11	-0.44	-0.24	0.02	-0.03	-0.04	-0.40	
P09	-0.50	-0.30	-0.47	0.41	-0.43	0.38	-0.47	-0.98
P10					-0.01	0.15		
P11	-0.13	-1.18		-0.70	-7.29		0.43	0.27
P12	0.06	-0.87	0.14	1.11	0.36	-1.04	-0.45	-1.26
P13		-8.99	0.58	0.02	0.74	-0.38	-2.93	1.39
P14		0.87	0.90		-0.74			
P15	-2.92	-2.56		0.19	0.87	-0.01	-1.63	-2.60
P16		0.04						
P17	0.00	-0.11	0.16	-0.13	0.21	1.20	0.42	0.51
P18	-0.09	-0.83	-0.15	0.05	-0.07			
P19					0.53	0.88		
P20					-0.06			
P21	-0.16	0.08	0.50	-0.20	0.49	-0.80	0.46	0.08
P22	0.37	-0.56	0.37	-0.14	0.04	0.09		
P23	-0.06	-0.26	0.49	-0.04		-0.23	0.51	0.38
P24	0.18	-0.68	-0.28	-0.01	0.15	0.53	-1.30	-0.96
P25	0.31		-1.71	0.11			-1.72	
P26	-0.20	0.00	0.52	-0.25	0.32	-0.76	0.50	0.16
P27	-0.42	0.42		0.56	0.00	0.03	-0.38	-0.53

For the laboratories submitting estimates of uncertainty for their measurements, the corresponding E_n -numbers show fewer satisfactory result to those for **z**-scores.

For participants reporting unsatisfactory results, this suggests the presence of an undetected measurement bias or an underestimation of their reported uncertainties.

Excellent performances were received from participants **P03**, **P07**, **P08**, **P09**, **P18**, **P22**, **P23** and **P27**, each of whom submitted results for 4 or more measurands achieving 100% perfect score on the basis of both performance measures.

The outstanding laboratories in this round of the PT scheme were participants **P07** and **P09** with a 100% perfect score on the basis of both performance measures for all **eight** measurands.

25/0642-0648/R Page **14** of **33**

Annex A - Detailed results by measurand

Detailed results for all measurands in all mixtures are shown in subsequent charts.

In each chart, the reported results are shown with the dots in terms of a relative difference (in percent) from the assigned reference value. The reported uncertainties (where supplied) are shown as "error bars" on the reported values.

In each chart the bound limit lines surrounding the zero relative difference signify

- the percentage relative uncertainty on the reference value, $\%U(x_{ref})$ k=2 (in blue)
- the |z|=2 satisfactory limit (in green)
- the |z|=3 unsatisfactory limit (in red)

This annex also includes additional statistics presenting consensus values from the pool of laboratories on the basis of raw data and correct data (following the removal of outlying reported values).

Additional tables also show repeatability standard deviation (s_r), between laboratory standard deviation (s_L) and reproducibility standard deviation (s_R) on the basis of raw and corrected data. The data has been calculated in accordance with the robust statistical methods in ISO 5725 Parts 1 and 2. The detailed calculations made to derive these results are outside the scope of this report but will be provided to participants on request from the scheme coordinator.

25/0642-0648/R Page **15** of **33**

Measurand/ Mixture

sulphur dioxide

Reference

X _{ref}	$U(x_{ref}) k=2$	_	σ	_
115.6	1.2	μmol/mol	2.9	μmol/mol

Reported data


reported data					
id	value (µmol/mol)	U (k=2) (μmol/mol)	relative difference	z-score	E _n -number
P01	116.4	1.0	0.65%	0.26	0.47
P02					
P03					
P04					
P05					
P06	118.9	0.9	2.83%	1.13	2.21
P07	117.1	7.3	1.28%	0.51	0.20
P08	116.4	6.8	0.67%	0.27	0.11
P09	114.0	3.0	-1.37%	-0.55	-0.50
P10					
P11	114.7	6.8	-0.78%	-0.31	-0.13
P12	116.1	7.8	0.42%	0.17	0.06
P13					
P14					
P15	110.1	1.4	-4.72%	-1.89	-2.92
P16					
P17	115.6	6.2	-0.01%	-0.01	0.00
P18	115.0	7.0	-0.56%	-0.22	-0.09
P19					
P20					
P21	114.4	7.3	-1.04%	-0.41	-0.16
P22	117.7	5.4	1.79%	0.72	0.37
P23	115.3	5.8	-0.30%	-0.12	-0.06
P24	117.0	7.5	1.19%	0.47	0.18
P25	122.0	20.7	5.54%	2.21	0.31
P26	114.2	7.3	-1.25%	-0.50	-0.20
P27	114.8	1.5	-0.70%	-0.28	-0.42

25/0642-0648/R Page **16** of **33**

m	115.5	
S _r	0.3	0.26%
S _L	2.0	1.76%
S _R	2.1	1.78%
р	17	

Consensus values (corrected)

m	115.8	
S _r	0.3	0.27%
S _L	1.4	1.22%
S _R	1.4	1.25%
р	15	



25/0642-0648/R Page **17** of **33**

Measurand/ Mixture	propane				
Reference	x _{ref}	$U(x_{ref}) k=2$	-	σ	'
	27.14	0.26	µmol/mol	1.36	µmol/mol
Reported data					•
id	value (µmol/mol)	U (k=2) (µmol/mol)	relative difference	z -score	E _n -number
P01	27.10	0.87	-0.15%	-0.03	-0.05
P02					
P03					
P04					
P05					
P06	27.34	0.87	0.72%	0.14	0.21
P07	26.94	2.15	-0.74%	-0.15	-0.09
P08	26.41	1.64	-2.69%	-0.54	-0.44
P09	26.44	2.35	-2.58%	-0.52	-0.30
P10					
P11	25.60	1.28	-5.67%	-1.13	-1.18
P12	25.89	1.42	-4.62%	-0.92	-0.87
P13	23.82	0.26	-12.23%	-2.45	-8.99
P14	27.47	0.27	1.22%	0.24	0.87
P15	24.78	0.89	-8.70%	-1.74	-2.56
P16	27.28	3.55	0.52%	0.10	0.04
P17	26.83	2.95	-1.15%	-0.23	-0.11
P18	26.59	0.60	-2.01%	-0.40	-0.83
P19					
P20					
P21	27.35	2.72	0.77%	0.15	0.08
P22	26.49	1.14	-2.41%	-0.48	-0.56
P23	26.52	2.33	-2.29%	-0.46	-0.26
P24	25.40	2.55	-6.41%	-1.28	-0.68
P25					

2.67

0.87

27.14

27.53

P26

P27

25/0642-0648/R Page **18** of **33**

0.01%

1.42%

0.00

0.28


0.00

0.42

m	26.38	
s _r	0.22	0.85%
S _L	1.00	3.78%
S _R	1.02	3.88%
р	19	

Consensus values (corrected)

m	26.54	
S _r	0.23	0.87%
S _L	0.78	2.92%
S _R	0.81	3.05%
р	18	



25/0642-0648/R Page **19** of **33**

Measurand/ Mixture	nitric oxide				
Reference	X _{ref}	$U(x_{ref}) k=2$	_	σ	
	82.45	0.46	μmol/mol	2.06	µmol/mol
Reported data			_		
id	value (µmol/mol)	U (k=2) (µmol/mol)	relative difference	z -score	E _n -number
P01					
P02					
P03					
P04					
P05	83.57	2.73	1.36%	0.54	0.41
P06					
P07	82.58	6.61	0.16%	0.06	0.02
P08	80.92	6.30	-1.86%	-0.74	-0.24
P09	81.14	2.73	-1.59%	-0.64	-0.47
P10					
P11					
P12	82.79	2.40	0.41%	0.16	0.14
P13	83.22	1.25	0.93%	0.37	0.58
P14	83.31	0.83	1.04%	0.42	0.90
P15					
P16					
P17	83.11	4.07	0.81%	0.32	0.16
P18	81.94	3.40	-0.61%	-0.25	-0.15
P19					
P20					
P21	84.69	4.41	2.72%	1.09	0.50
P22	83.54	2.93	1.32%	0.53	0.37
P23	84.25	3.62	2.19%	0.87	0.49
P24	81.26	4.27	-1.44%	-0.58	-0.28
P25	70.00	7.27	-15.10%	-6.04	-1.71
P26	84.79	4.44	2.84%	1.14	0.52

P27


25/0642-0648/R Page **20** of **33**

m	82.82	
S _r	0.22	0.26%
S _L	1.77	2.14%
S _R	1.79	2.16%
р	15	

Consensus values (corrected)

m	82.93	
s _r	0.22	0.26%
S _L	1.32	1.59%
S _R	1.34	1.61%
р	14	





25/0642-0648/R Page **21** of **33**

Measurand/ Mixture

carbon monoxide

-				
ĸ	Δt	Or	en	CO
1/	CI	CI	CII	L

X _r	ef U(x _r	$_{\text{ef}}$) $k=2$:
23!	5.8	1.9 μmo	I/mol 3.	5 μmol/mol

Reported data


id value (μmol/mol) U (k=2) (μmol/mol) relative difference z-score E _n -number P01 238.8 1.9 1.28% 0.85 1.12 P02						
P02 P03 236.5 23.6 0.29% 0.19 0.03 P04 P05 P06 240.2 2.8 1.87% 1.25 1.29 P07 238.5 10.7 1.15% 0.77 0.25 P08 236.1 13.7 0.14% 0.10 0.02 P09 237.5 3.8 0.73% 0.49 0.41 P10 P11 227.8 11.2 -3.38% -2.25 -0.70 P12 244.0 7.1 3.46% 2.30 1.11 P13 235.9 3.8 0.04% 0.03 0.02 P14 P15 236.3 1.5 0.20% 0.13 0.19 P16 P17 234.2 12.4 -0.68% -0.45 -0.13 P18 236.4 10.7 0.25% 0.17 0.05 P19 P20 P21 233.0 13.5 -1.18% -0.79 -0.20 P22 235.2 3.7 -0.25% -0.17 -0.14 P23 235.3 11.5 -0.22% -0.14 -0.04 P24 235.6 13.5 -0.07% -0.05 -0.01 P25 239.1 30.6 1.40% 0.93 0.11 P26 232.5 13.1 -1.41% -0.94 -0.25	id		. ,		z -score	E _n -number
P03 236.5 23.6 0.29% 0.19 0.03 P04 P05 P06 240.2 2.8 1.87% 1.25 1.29 P07 238.5 10.7 1.15% 0.77 0.25 P08 236.1 13.7 0.14% 0.10 0.02 P09 237.5 3.8 0.73% 0.49 0.41 P10 P11 227.8 11.2 -3.38% -2.25 -0.70 P12 244.0 7.1 3.46% 2.30 1.11 P13 235.9 3.8 0.04% 0.03 0.02 P14 P15 236.3 1.5 0.20% 0.13 0.19 P16 P17 234.2 12.4 -0.68% -0.45 -0.13 P18 236.4 10.7 0.25% 0.17 0.05 P19 P20 P21 233.0 13.5 -1.18% -0.79 -0.20 P22 235.2 3.7 -0.25% -0.17 -0.14 P23 235.6 13.5 -0.22% -0.14 -0.04 P24 235.6 13.5 -0.07% -0.05 -0.01 P25 239.1 30.6 1.40% 0.93 0.11 P26 232.5 13.1 -1.41% -0.94 -0.25	P01	238.8	1.9	1.28%	0.85	1.12
P04 P05 P06	P02					
P05 P06	P03	236.5	23.6	0.29%	0.19	0.03
P06 240.2 2.8 1.87% 1.25 1.29 P07 238.5 10.7 1.15% 0.77 0.25 P08 236.1 13.7 0.14% 0.10 0.02 P09 237.5 3.8 0.73% 0.49 0.41 P10 P10 P11 227.8 11.2 -3.38% -2.25 -0.70 P12 244.0 7.1 3.46% 2.30 1.11 P13 235.9 3.8 0.04% 0.03 0.02 P14 P15 236.3 1.5 0.20% 0.13 0.19 P16 P17 234.2 12.4 -0.68% -0.45 -0.13 P18 236.4 10.7 0.25% 0.17 0.05 P19 P20 P21 233.0 13.5 -1.18% -0.79 -0.20 P22 235.2 3.7 -0.25% -0.17 -0.14 P23 235.3 11.5 -0.22% -0.14 -0.04 <	P04					
P07 238.5 10.7 1.15% 0.77 0.25 P08 236.1 13.7 0.14% 0.10 0.02 P09 237.5 3.8 0.73% 0.49 0.41 P10 P10 P11 227.8 11.2 -3.38% -2.25 -0.70 P12 244.0 7.1 3.46% 2.30 1.11 P13 235.9 3.8 0.04% 0.03 0.02 P14 P15 236.3 1.5 0.20% 0.13 0.19 P16 P17 234.2 12.4 -0.68% -0.45 -0.13 P18 236.4 10.7 0.25% 0.17 0.05 P19 P20 P21 233.0 13.5 -1.18% -0.79 -0.20 P22 235.2 3.7 -0.25% -0.17 -0.14 P23 235.3 11.5 -0.22% -0.14 -0.04 P24 235.6 13.5 -0.07% -0.05 -0.01	P05					
P08 236.1 13.7 0.14% 0.10 0.02 P09 237.5 3.8 0.73% 0.49 0.41 P10 P11 227.8 11.2 -3.38% -2.25 -0.70 P12 244.0 7.1 3.46% 2.30 1.11 P13 235.9 3.8 0.04% 0.03 0.02 P14 P15 236.3 1.5 0.20% 0.13 0.19 P16 P17 234.2 12.4 -0.68% -0.45 -0.13 P18 236.4 10.7 0.25% 0.17 0.05 P19 P20 P21 233.0 13.5 -1.18% -0.79 -0.20 P22 235.2 3.7 -0.25% -0.17 -0.14 P23 235.6 13.5 -0.22% -0.14 -0.04 P24 235.6 13.5 -0.07% -0.05 -0.01 P25 239.1 30.6 1.40% 0.93 0.11 P26 232.5 13.1 -1.41% -0.94 -0.25	P06	240.2	2.8	1.87%	1.25	1.29
P09 237.5 3.8 0.73% 0.49 0.41 P10 P11 227.8 11.2 -3.38% -2.25 -0.70 P12 244.0 7.1 3.46% 2.30 1.11 P13 235.9 3.8 0.04% 0.03 0.02 P14 P15 236.3 1.5 0.20% 0.13 0.19 P16 P17 234.2 12.4 -0.68% -0.45 -0.13 P18 236.4 10.7 0.25% 0.17 0.05 P19 P20 P21 233.0 13.5 -1.18% -0.79 -0.20 P22 235.2 3.7 -0.25% -0.17 -0.14 P23 235.3 11.5 -0.22% -0.14 -0.04 P24 235.6 13.5 -0.07% -0.05 -0.01 P25 239.1 30.6 1.40% 0.93 0.11 P26 232.5 13.1 -1.41% -0.94 -0.25	P07	238.5	10.7	1.15%	0.77	0.25
P10 P11	P08	236.1	13.7	0.14%	0.10	0.02
P11 227.8 11.2 -3.38% -2.25 -0.70 P12 244.0 7.1 3.46% 2.30 1.11 P13 235.9 3.8 0.04% 0.03 0.02 P14 P15 236.3 1.5 0.20% 0.13 0.19 P16 P17 234.2 12.4 -0.68% -0.45 -0.13 P18 236.4 10.7 0.25% 0.17 0.05 P19 P20 P21 233.0 13.5 -1.18% -0.79 -0.20 P22 235.2 3.7 -0.25% -0.17 -0.14 P23 235.3 11.5 -0.22% -0.14 -0.04 P24 235.6 13.5 -0.07% -0.05 -0.01 P25 239.1 30.6 1.40% 0.93 0.11 P26 232.5 13.1 -1.41% -0.94 -0.25	P09	237.5	3.8	0.73%	0.49	0.41
P12 244.0 7.1 3.46% 2.30 1.11 P13 235.9 3.8 0.04% 0.03 0.02 P14 P15 236.3 1.5 0.20% 0.13 0.19 P16 P17 234.2 12.4 -0.68% -0.45 -0.13 P18 236.4 10.7 0.25% 0.17 0.05 P19 P20 P21 233.0 13.5 -1.18% -0.79 -0.20 P22 235.2 3.7 -0.25% -0.17 -0.14 P23 235.3 11.5 -0.22% -0.14 -0.04 P24 235.6 13.5 -0.07% -0.05 -0.01 P25 239.1 30.6 1.40% 0.93 0.11 P26 232.5 13.1 -1.41% -0.94 -0.25	P10					
P13 235.9 3.8 0.04% 0.03 0.02 P14 P15 236.3 1.5 0.20% 0.13 0.19 P16 P17 234.2 12.4 -0.68% -0.45 -0.13 P18 236.4 10.7 0.25% 0.17 0.05 P19 P20 P21 233.0 13.5 -1.18% -0.79 -0.20 P22 235.2 3.7 -0.25% -0.17 -0.14 P23 235.3 11.5 -0.22% -0.14 -0.04 P24 235.6 13.5 -0.07% -0.05 -0.01 P25 239.1 30.6 1.40% 0.93 0.11 P26 232.5 13.1 -1.41% -0.94 -0.25	P11	227.8	11.2	-3.38%	-2.25	-0.70
P14 P15 236.3 1.5 0.20% 0.13 0.19 P16 P17 234.2 12.4 -0.68% -0.45 -0.13 P18 236.4 10.7 0.25% 0.17 0.05 P19 P20 P21 233.0 13.5 -1.18% -0.79 -0.20 P22 235.2 3.7 -0.25% -0.17 -0.14 P23 235.3 11.5 -0.22% -0.14 -0.04 P24 235.6 13.5 -0.07% -0.05 -0.01 P25 239.1 30.6 1.40% 0.93 0.11 P26 232.5 13.1 -1.41% -0.94 -0.25	P12	244.0	7.1	3.46%	2.30	1.11
P15 236.3 1.5 0.20% 0.13 0.19 P16 P17 234.2 12.4 -0.68% -0.45 -0.13 P18 236.4 10.7 0.25% 0.17 0.05 P19 P20 P21 233.0 13.5 -1.18% -0.79 -0.20 P22 235.2 3.7 -0.25% -0.17 -0.14 P23 235.3 11.5 -0.22% -0.14 -0.04 P24 235.6 13.5 -0.07% -0.05 -0.01 P25 239.1 30.6 1.40% 0.93 0.11 P26 232.5 13.1 -1.41% -0.94 -0.25	P13	235.9	3.8	0.04%	0.03	0.02
P16 P17 234.2 12.4 -0.68% -0.45 -0.13 P18 236.4 10.7 0.25% 0.17 0.05 P19 P20 P21 233.0 13.5 -1.18% -0.79 -0.20 P22 235.2 3.7 -0.25% -0.17 -0.14 P23 235.3 11.5 -0.22% -0.14 -0.04 P24 235.6 13.5 -0.07% -0.05 -0.01 P25 239.1 30.6 1.40% 0.93 0.11 P26 232.5 13.1 -1.41% -0.94 -0.25	P14					
P17 234.2 12.4 -0.68% -0.45 -0.13 P18 236.4 10.7 0.25% 0.17 0.05 P19 P20 P21 233.0 13.5 -1.18% -0.79 -0.20 P22 235.2 3.7 -0.25% -0.17 -0.14 P23 235.3 11.5 -0.22% -0.14 -0.04 P24 235.6 13.5 -0.07% -0.05 -0.01 P25 239.1 30.6 1.40% 0.93 0.11 P26 232.5 13.1 -1.41% -0.94 -0.25	P15	236.3	1.5	0.20%	0.13	0.19
P18 236.4 10.7 0.25% 0.17 0.05 P19 0.05 0.17 0.05 0.17 0.05 0.17 0.20 0.20 0.20 0.20 0.25% 0.17 0.14 0.14 0.14 0.04 0.04 0.04 0.04 0.05 0.01 0.01 0.05 0.01 0.01 0.93 0.11 0.93 0.11 0.94 0.94 0.025 0.05 0.025 0.05 0.025 0.01 0.02	P16					
P19 P20 P21 233.0 13.5 -1.18% -0.79 -0.20 P22 235.2 3.7 -0.25% -0.17 -0.14 P23 235.3 11.5 -0.22% -0.14 -0.04 P24 235.6 13.5 -0.07% -0.05 -0.01 P25 239.1 30.6 1.40% 0.93 0.11 P26 232.5 13.1 -1.41% -0.94 -0.25	P17	234.2	12.4	-0.68%	-0.45	-0.13
P20 P21 233.0 13.5 -1.18% -0.79 -0.20 P22 235.2 3.7 -0.25% -0.17 -0.14 P23 235.3 11.5 -0.22% -0.14 -0.04 P24 235.6 13.5 -0.07% -0.05 -0.01 P25 239.1 30.6 1.40% 0.93 0.11 P26 232.5 13.1 -1.41% -0.94 -0.25	P18	236.4	10.7	0.25%	0.17	0.05
P21 233.0 13.5 -1.18% -0.79 -0.20 P22 235.2 3.7 -0.25% -0.17 -0.14 P23 235.3 11.5 -0.22% -0.14 -0.04 P24 235.6 13.5 -0.07% -0.05 -0.01 P25 239.1 30.6 1.40% 0.93 0.11 P26 232.5 13.1 -1.41% -0.94 -0.25	P19					
P22 235.2 3.7 -0.25% -0.17 -0.14 P23 235.3 11.5 -0.22% -0.14 -0.04 P24 235.6 13.5 -0.07% -0.05 -0.01 P25 239.1 30.6 1.40% 0.93 0.11 P26 232.5 13.1 -1.41% -0.94 -0.25	P20					
P23 235.3 11.5 -0.22% -0.14 -0.04 P24 235.6 13.5 -0.07% -0.05 -0.01 P25 239.1 30.6 1.40% 0.93 0.11 P26 232.5 13.1 -1.41% -0.94 -0.25	P21	233.0	13.5	-1.18%	-0.79	-0.20
P24 235.6 13.5 -0.07% -0.05 -0.01 P25 239.1 30.6 1.40% 0.93 0.11 P26 232.5 13.1 -1.41% -0.94 -0.25	P22	235.2	3.7	-0.25%	-0.17	-0.14
P25 239.1 30.6 1.40% 0.93 0.11 P26 232.5 13.1 -1.41% -0.94 -0.25	P23	235.3	11.5	-0.22%	-0.14	-0.04
P26 232.5 13.1 -1.41% -0.94 -0.25	P24	235.6	13.5	-0.07%	-0.05	-0.01
	P25	239.1	30.6	1.40%	0.93	0.11
P27 240.5 8.2 2.00% 1.33 0.56	P26	232.5	13.1	-1.41%	-0.94	-0.25
	P27	240.5	8.2	2.00%	1.33	0.56

25/0642-0648/R Page **22** of **33**

m	236.3	
S _r	0.3	0.14%
S _L	3.5	1.49%
S _R	3.5	1.50%
р	19	

Consensus values (corrected)

m	236.4	
s _r	0.4	0.15%
S _L	2.3	0.97%
S _R	2.3	0.98%
р	17	



25/0642-0648/R Page **23** of **33**

			-		
Measurand/ Mixture	oxygen				
Reference	X _{ref}	$U(x_{ref}) k=2$		σ	
	6.827	0.037	%mol/mol	0.078	%mol/mol
Reported data			_		•
id	value (%mol/mol)	U (k=2) (%mol/mol)	relative difference	z -score	E _n -number
P01	6.850	0.144	0.34%	0.29	0.15
P02					
P03	6.835	0.684	0.12%	0.11	0.01
P04	6.815	0.130	-0.18%	-0.15	-0.09
P05	6.822	0.015	-0.07%	-0.06	-0.12
P06	6.882	0.015	0.81%	0.71	1.38
P07	6.832	0.290	0.07%	0.06	0.02
P08	6.817	0.290	-0.15%	-0.13	-0.03
P09	6.809	0.019	-0.26%	-0.23	-0.43
P10	6.820	0.500	-0.10%	-0.09	-0.01
P11	6.257	0.069	-8.35%	-7.28	-7.29
P12	6.892	0.179	0.95%	0.83	0.36
P13	6.885	0.069	0.85%	0.74	0.74
P14	6.770	0.068	-0.83%	-0.73	-0.74
P15	7.050	0.255	3.27%	2.85	0.87
P16					
P17	6.900	0.350	1.07%	0.93	0.21
P18	6.818	0.130	-0.13%	-0.11	-0.07
P19	6.882	0.096	0.80%	0.70	0.53
P20	6.815	0.210	-0.18%	-0.15	-0.06
P21	6.904	0.154	1.13%	0.98	0.49
P22	6.836	0.232	0.13%	0.11	0.04
P23					
P24	6.851	0.153	0.35%	0.31	0.15
P25	6.860		0.48%	0.42	
P26	6.878	0.153	0.75%	0.65	0.32

P27

6.826

0.201

25/0642-0648/R Page **24** of **33**

-0.01%


-0.01

0.00

m	6.830	
s _r	0.046	0.67%
S _L	0.141	2.06%
S _R	0.148	2.17%
р	24	

Consensus values (corrected)

m	6.858	
s _r	0.015	0.21%
S _L	0.055	0.81%
S _R	0.057	0.83%
р	23	



Page **25** of **33** 25/0642-0648/R

Measurand/ carbon dioxide **Mixture** Reference $U(x_{ref}) k=2$ $\mathbf{x}_{\mathsf{ref}}$ σ 6.164 0.025 0.072 %mol/mol %mol/mol Reported data U (k=2) relative value id z-score $\mathbf{E}_{\mathbf{n}}$ -number (%mol/mol) (%mol/mol) difference P01 6.137 0.266 -0.44% -0.38 -0.10 P02 P03 0.624 0.13 6.244 1.29% 1.11 P04 6.265 0.270 1.64% 1.41 0.37 P05 0.013 6.057 -1.74% -1.49 -3.81 P06 6.161 0.016 -0.05% -0.04 -0.11 P07 6.167 0.270 0.06% 0.05 0.01 P08 6.142 0.520 -0.36% -0.31 -0.04 P09 6.175 0.014 0.18% 0.15 0.38

0.500

0.156

0.055

0.267

0.280

0.069

0.207

0.223

0.341

0.215

0.206

0.268

1.23%

-2.68%

-0.37%

-0.06%

5.45%

1.04%

-2.69%

0.34%

-1.27%

1.87%

-2.34%

-2.56%

0.11%

1.06

-2.30

-0.32

-0.05

4.69

0.90

0.29

-1.09

1.61

-2.21

0.10

0.15

-1.04

-0.38

-0.01

1.20

0.88

-0.80

0.09

-0.23

0.53

-0.76

0.03

P10

P11 P12

P13

P14 P15

P16 P17

P18 P19

P20

P21

P22

P23

P24

P25

P26

P27

6.240

5.999

6.141

6.161

6.500

6.228

5.998

6.185

6.086

6.279

6.020

6.006


6.171

25/0642-0648/R Page **26** of **33**

т	6.164	
S _r	0.019	0.31%
S _L	0.117	1.90%
S _R	0.119	1.93%
р	21	

Consensus values (corrected)

m	6.145	
s _r	0.020	0.32%
S _L	0.087	1.42%
S _R	0.090	1.46%
р	20	



Page **27** of **33** 25/0642-0648/R

Measurand/ Mixture nitric oxide (NO/NO2 mix)

Reference

x_{ref}	$U(x_{ref}) k=2$	_	σ	_
86.57	0.48	μmol/mol	2.16	μmol/mol

Reported data


	id	value (μmol/mol)	U (k=2) (μmol/mol)	relative difference	z -s core	E _n -number
1	P01	87.47	0.44	1.04%	0.42	1.39
	P02	76.86	1.54	-11.21%	-4.49	-6.03
	P03	86.68	8.67	0.12%	0.05	0.01
	P04					
	P05					
	P06	86.18	1.57	-0.45%	-0.18	-0.24
	P07	86.54	6.92	-0.03%	-0.01	0.00
	P08	84.02	6.30	-2.95%	-1.18	-0.40
	P09	84.89	3.52	-1.94%	-0.77	-0.47
	P10					
	P11	88.01	3.34	1.66%	0.67	0.43
	P12	85.44	2.48	-1.31%	-0.52	-0.45
	P13	82.67	1.24	-4.51%	-1.80	-2.93
	P14					
	P15	82.48	2.47	-4.72%	-1.89	-1.63
	P16					
	P17	88.42	4.33	2.14%	0.85	0.42
	P18					
	P19					
	P20					
	P21	88.69	4.53	2.44%	0.98	0.46
	P22					
	P23	88.53	3.81	2.26%	0.90	0.51
	P24	81.01	4.26	-6.42%	-2.57	-1.30
	P25	73.40	7.62	-15.21%	-6.09	-1.72
	P26	88.83	4.52	2.61%	1.04	0.50
	P27	83.81	7.29	-3.19%	-1.28	-0.38

25/0642-0648/R Page **28** of **33**

m	85.52	
s _r	0.17	0.20%
S _L	3.05	3.56%
S _R	3.05	3.57%
р	18	

Consensus values (corrected)

m	85.83	
s _r	0.17	0.20%
S _L	2.53	2.95%
S _R	2.54	2.96%
р	16	



25/0642-0648/R Page **29** of **33**

Measurand/ Mixture nitrogen oxides (NO/NO2 mix)

Reference

\mathbf{x}_{ref}	$U(x_{ref}) k=2$	_	σ	
97.77	0.69	μmol/mol	2.44	μmol/mol

Reported data


E _n -number	z -score	relative difference	U (k=2) (μmol/mol)	value (µmol/mol)	id
-2.17	-0.74	-1.86%	0.48	95.95	P01
-12.07	-8.37	-20.91%	1.55	77.32	P01
	-0.05	-20.91%	9.76	97.64	P02
-0.01	-0.05	-0.14%	9.76	97.04	
					P04
4.54	4.47	2.020/	4.72	04.04	P05
-1.54	-1.17	-2.93%	1.72	94.91	P06
-0.20	-0.64	-1.59%	7.69	96.21	P07
	-1.91	-4.78%		93.10	P08
-0.98	-1.55	-3.88%	3.80	93.98	P09
					P10
0.27	0.46	1.15%	4.15	98.89	P11
-1.26	-1.72	-4.31%	3.27	93.56	P12
1.39	1.49	3.72%	2.54	101.41	P13
					P14
-2.60	-2.97	-7.43%	2.71	90.50	P15
					P16
0.51	1.04	2.59%	4.91	100.30	P17
					P18
					P19
					P20
0.08	0.21	0.53%	6.30	98.29	P21
					P22
0.38	0.68	1.69%	4.28	99.42	P23
-0.96	-2.39	-5.97%	6.04	91.93	P24
	-5.80	-14.49%		83.60	P25
0.16	0.42	1.04%	6.29	98.79	P26
-0.53	-1.78	-4.45%	8.13	93.42	P27

25/0642-0648/R Page **30** of **33**

m	95.59	
s _r	0.29	0.31%
S _L	4.53	4.74%
S _R	4.54	4.75%
р	18	

Consensus values (corrected)

т	96.06	
S _r	0.29	0.30%
S _L	3.43	3.57%
S _R	3.44	3.58%
р	17	



25/0642-0648/R Page **31** of **33**

Annex B - Converter efficiency

If the reported nitric oxide (NO) measurement of the NO/NO_2 mixture, for each participant, is subtracted from that of their reported nitrogen oxides (NO_X) result, then the nitrogen dioxide (NO_2) result from their measurements can be directly calculated. This derived NO_2 measurement result can be used to calculate the converter efficiency of their analyser where appropriate.

The table below gives the derived results for nitrogen dioxide and the calculated converter efficiencies for each reporting participant. Their uncertainties have been calculated by adding the uncertainties of their NO and NO_X reported results in quadrature.

25/0642-0648/R Page **32** of **33**

			_		
Reference	X _{ref}	$U(x_{ref}) k=2$			
	11.20	0.84	µmol/mol		
Reported data	a				
id	value	U (k=2)	difference	converter	
	(µmol/mol)	(µmol/mol)	(µmol/mol)	efficiency (%)	E _n -number
P01	8.48	0.65	-2.7	75.7%	-2.56
P02	0.46	2.18	-10.7	4.1%	-4.60
P03	10.96	13.06	-0.2	97.9%	-0.02
P04					
P05					
P06	8.73	2.33	-2.5	77.9%	-1.00
P07	9.67	10.35	-1.5	86.3%	-0.15
P08	9.08	6.30	-2.1	81.1%	-0.33
P09	9.09	5.18	-2.1	81.1%	-0.40
P10					
P11	10.88	5.33	-0.3	97.1%	-0.06
P12	8.12	4.11	-3.1	72.5 %	-0.73
P13	18.74	2.82	7.5	167.3%	2.56
P14					
P15	8.02	3.66	-3.2	71.6%	-0.85
P16					
P17	11.88	6.55	0.7	106.1%	0.10
P18					
P19					
P20					
P21	9.60	7.76	-1.6	85.7%	-0.20
P22					
P23	10.90	5.72	-0.3	97.3%	-0.05
P24	10.92	7.39	-0.3	97.5%	-0.04
P25	10.20	7.62	-1.0	91.1%	-0.13
P26	9.96	7.75	-1.2	88.9%	-0.16

10.92

9.61

P27

For appropriate measurement of nitrogen dioxide by the conversion of NO_2 to NO using a converter and subsequent measurement by chemiluminescence, the efficiency of the converter should be above 95% (in accordance with BS EN 14792).

85.8%

-1.6

-0.15

25/0642-0648/R Page **33** of **33**